

HIGH OR LOW VALUES

Control unit console displays “high” value:



The measured pressure is higher than 475 mbar. For safety reason, the pump automatically stops, and the “high” value is displayed.

Control unit console displays “low” value:



The measured pressure is lower than 25 mbar. Try to perform the measurement again, taking care that the tubes are not kinked at any location. If the “low” value is displayed again, it means that the cervix is extremely soft.